



## WI State Laboratory of Hygiene *Cryptosporidium* Proficiency Testing Program

May 12, 2016

Dr. Yeggie Dearborn  
82 Mary St.  
Suite 2  
San Francisco, CA 94103

**Report #: CA00001-1623.1-201604-01**

Dear Dr. Dearborn:

Thank you for participating in the April 2016 round of the *Cryptosporidium* Proficiency Testing (PT) Program. You are receiving two reports for this round. This first report is specific to your laboratory's performance using Method 1623.1 and includes this letter, a table showing your laboratory's recovery values, and a chart illustrating your laboratory's mean recovery by PT round. The second report, also included in this correspondence, includes the PT General Summary Report.

The acceptable mean percent recovery range for the April 2016 round was 35.7 - 89.7 %. **Your laboratory's mean recovery percentage falls within this acceptable range; therefore, your laboratory has passed the April 2016 PT event.**

Your proficiency testing results are confidential. They will not be released without written authorization from your laboratory. The general summary of participant performance may be disclosed however, any unique identifiers will be removed.

Please note that reports are not to be used as an endorsement of your laboratory by the University of Wisconsin or the Wisconsin State Laboratory of Hygiene.

If you have any questions, please contact Becky Hoffman or Marty Collins at the PT program email: [CryptoPT@slh.wisc.edu](mailto:CryptoPT@slh.wisc.edu) or (608) 224-6260.

Sincerely,

A handwritten signature in black ink that reads "Becky Hoffman".

Becky Hoffman  
WSLH *Cryptosporidium* PT Program Coordinator